



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Dynatek Labs, Inc.***  
105 E. 4<sup>th</sup> Street, Galena, MO 65656

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Mechanical Testing***  
*(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President/Operations Manager

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*Initial Accreditation Date:*

April 15, 2020

*Issue Date:*

April 15, 2020

*Expiration Date:*

July 31, 2022

*Accreditation No.:*

102435

*Certificate No.:*

L20-203

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjilabs.com](http://www.pjilabs.com)*



# Certificate of Accreditation: Supplement

## Dynatek Labs, Inc.

105 E. 4<sup>th</sup> Street, Galena, MO 65656

Contact Name: Jessica Davis Phone: 417-357-6155

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical <sup>F</sup>	Medical Devices	Acute Particulate Matter Generation	ASTM F2743	5 µm to 900 µm
	Vascular Devices	Pulsatile Durability	ASTM F2477 ISO 25539-1 Clause 8.5.2.3, Annex D, D.5.2.3 ISO 25539-2 Clause 8.6.3.5.4; Annex D, D.5.3.10	1.5 mm to 50 mm
		Chronic Particulate Matter Generation with Radial Fatigue	ASTM F2477 ISO 25539-2 Clauses 8.6.3.5.4, 8.6.3.5.6; Annex D, D.5.3.10	5 µm to 900 µm (particulate) 1.5 mm to 50 mm (device size)
	Endovascular devices, Vascular prostheses, Mock vasculature	Determination of Dynamic Compliance	ISO 7198, Annex A.5.9	1.5 mm to 50 mm

1. The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this testing at its fixed location.