



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Dynatek Labs, Inc.
105 E. 4th Street, Galena, MO 65656

*(Hereinafter called the Organization) and hereby declares that Organization is accredited
in accordance with the recognized International Standard:*

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the
operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Mechanical Testing
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

April 15, 2020

Issue Date:

May 27, 2022

Expiration Date:

August 31, 2024

Accreditation No.:

102435

Certificate No.:

L22-390

*The validity of this certificate is maintained through ongoing assessments based on a
continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjilabs.com*



Certificate of Accreditation: Supplement

Dynatek Labs, Inc.

105 E. 4th Street, Galena, MO 65656
Contact Name: Jessica Davis Phone: 417-357-6155

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical ^F	Medical Devices	Acute Particulate Matter Generation	ASTM F2743 ISO 25539-2 Clause 8.5.1.7; Annex D, D.5.2.6.1	5 µm to 900 µm
	Vascular Devices	Pulsatile Durability	ASTM F2477 ISO 25539-2 Clause 8.5.2.3; Annex D, D.5.2.3 ISO 25539-2 Clause 8.5.1.5; Annex D, D.5.3.3	1.5 mm to 50 mm
		Chronic Particulate Matter Generation with Radial Fatigue	ASTM F2477 ISO 25539-2 Clauses 8.5.1.5; Annex D, D.5.3.3	5 µm to 900 µm (particulate) 1.5 mm to 50 mm (device size)
	Endovascular devices, Vascular prostheses, Mock vasculature	Determination of Dynamic Compliance	ISO 7198, Annex A.5.9	1.5 mm to 50 mm

1. The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this testing at its fixed location.